

Congratulations !

You have purchased the latest in Handheld Conductivity-TDS-pH-Temperature instrumentation. We trust that your new **WP-81** will give you many years of reliable service.

The **WP-81** is a breeze to operate. This manual has been designed to help you get started, and also contains some handy application tips. If at any stage you require assistance, please contact either your local TPS representative or the TPS factory in Brisbane.

The manual is divided into the following sections:

1. **Table of Contents**

Each major section of the handbook is clearly listed. Sub-sections have also been included to enable you to find the information you need at a glance.

2. **Introduction**

The introduction has a diagram and explanation of the display and controls of the **WP-81**. It also contains a full listing of all of the items that you should have received with your **WP-81**. Please take the time to read this section, as it explains some of items that are mentioned in subsequent sections.

3. **Main Section**

The main section of the handbook provides complete details of the **WP-81**, including operating modes, calibration, troubleshooting, specifications, and warranty terms.

4. **Appendices**

Appendices containing background information and application notes are provided at the back of this manual.

**WP-81 Cond, TDS,
pH, mV, Temp. Meter**

Version : 5.4

Date : 30-Aug-2005

Author : AB



TPS Pty Ltd
4 Jamberoo Street
Springwood, Brisbane,
Australia, 4127

Phone : (07) 32 900 400
International : 61 7 32 900 400

Fax : (07) 3808 4871
International : 61 7 3808 4871

E-mail : tps@tps.com.au

Web Site : www.tps.com.au


Contents

1. Introduction	5
1.1 WP-81 Display and Controls.....	5
1.2 Unpacking Information	7
1.3 Specifications.....	8
2. WP-81 Menu Structure	11
3. Operating Modes	12
3.1 Selecting Conductivity or TDS Mode	12
3.2 Selecting pH or Millivolts (mV) Mode.....	12
4. Conductivity Calibration.....	13
4.1 Calibration Procedure	13
4.2 Calibration Notes	15
4.3 Calibration Messages	15
5. TDS Calibration	16
5.1 Calibration Procedure	16
5.2 Calibration Notes	18
5.3 Calibration Messages	18
6. pH Calibration.....	19
6.1 Calibration Procedure	19
6.2 Calibration Notes	20
6.3 Calibration Messages	21
7. mV Calibration	21
8. Temperature Calibration.....	22
8.1 Calibration Procedure	22
8.2 Calibration Notes	22
8.3 Calibration Messages	23
8.4 Manual Temperature Setting	23
9. Good Laboratory Practices (GLP)	24
9.1 To recall GLP information on the display.....	24
9.2 Failed Calibration	26
9.3 Printing GLP Information to the RS232 Port.....	26
9.4 Instrument Serial Number.....	27
9.5 Additional GLP Features.....	27

10.	Notepad Function.....	28
10.1	Recording Readings into the Notepad.....	28
10.2	Recalling Records from the Notepad.....	28
10.3	Erasing Records from the Notepad	29
10.4	Printing Records from the Notepad to the RS232 Port.....	29
11.	Automatic Datalogging	30
12.	RS232 Port	32
12.1	Setting the Baud Rate	32
12.2	Sending Readings to the RS232 Port.....	32
12.3	RS232 Configuration.....	32
12.4	Communication and Statistical Software	32
12.5	Commands.....	33
12.6	Data Format	34
12.7	GLP Data Format.....	35
13.	Battery Saver Function	38
14.	Recharging the Battery	39
15.	Clock Function	40
15.1	Setting the Clock.....	40
15.2	Displaying or Hiding the Clock.....	40
16.	Selecting k=0.1 or k=10 Sensors	41
17.	Selecting Buffers for Auto Buffer Recognition	42
18.	Initialising the WP-81	43
19.	Instrument firmware version number.....	43
20.	Troubleshooting	44
20.1	General Errors.....	44
20.2	Conductivity and TDS Troubleshooting	45
20.3	pH Troubleshooting.....	47
20.4	Temperature Troubleshooting	48
21.	Appendices	49
21.1	Care, Cleaning and Maintenance of Conductivity Electrodes	49
21.2	Replatinising Conductivity Electrodes.....	50
21.3	pH Electrode Fundamentals	51
21.4	Checking the reference junction of a pH electrode.....	53
21.5	Determining if an instrument or pH sensor is faulty	54
22.	Warranty.....	55

1. Introduction

1.1 WP-81 Display and Controls



① **F1**

Press to record readings into memory. See section 10.1.

Also used to select pH buffers for automatic buffer recognition at pH calibration. See section 18.

② **F2**

Press to show or hide the date/time or temperature. See section 16.2.

Also used to select k=0.1 or k=10 sensor, when standard k=1 sensor is not being used.

③ **F3**

Press to start or stop automatic logging. See section 11.



Alternatively, press to transmit current reading plus date and time to the RS232 port. See section 12.2.



④ **F4**


Only used within the menu system on the **WP-81**.

⑤ **Menu**

Press to access the user-friendly menu system which makes the **WP-81** a breeze to operate.

⑥  and ⑧ 

The  and  keys are used for calibrating temperature readout (section 8.1), setting the manual temperature compensation (section 8.4), setting the clock (section 16.1), setting the automatic logging period (section 11), and displaying GLP information (section 9.1).

The  key is also used to initialise the **WP-81** at turn-on. See section 19.

+ 

Switches the **WP-81** on and off.

⑨ **Display**

32 character alpha-numeric display with user-friendly menu and prompting system. Shows Conductivity/TDS, pH and Temperature simultaneously. Date and time can also be displayed.

1.2 Unpacking Information

Before using your new **WP-81**, please check that the following accessories have been included:

	Part No
1. WP-81 Conductivity-TDS-pH-Temp. Instrument.....	121132
2. k=1/ATC/Temperature Sensor, 1m cable	122201
3. Combination pH Sensor.....	121207
4. 2.76mS/cm Conductivity Standard, 200mL.....	122306
5. 2 ppK TDS Standard, 200mL.....	122307
6. pH6.88 Buffer, 200mL.....	121306
7. pH4.00 Buffer, 200mL.....	121381
8. Battery charger	130037
9. WP-81 Handbook	130050

*Options that may have been ordered with your **WP-81**:*

1. k=1/ATC/Temperature Sensor, 5m cable	122198
2. k=10/ATC/Temperature Sensor, 5m cable	122220
3. k=0.1/ATC/Temperature Sensor, 1m cable	122229
4. Extended cable, order by the metre..... (up to 10m max total length)	130040
5. RS232 Serial Interface Cable	130041
6. Communication software for Windows 95 and later	130086
7. USB to Serial Adaptor (requires 130041 also).....	130087
8. Hard Carry Case	130059
9. Battery charger lead for 12V cigarette lighter socket.....	130046
10. Battery charger lead for 12V DC, with battery clips.....	130052
11. Solar Panel.....	130012

Other spares:

1. 6V NiMH Battery	130038
--------------------------	--------

1.3 Specifications

Conductivity

Ranges4 ranges, with automatic range selection.
k=0.1 Sensor ...2.000 uS/cm to 2000 uS/cm
k=1.0 Sensor ...20.00 uS/cm to 20.00 mS/cm
k=10 Sensor200.0 uS/cm to 200.0 mS/cm

Resolution.....0.05% of selected range

Accuracy..... $\pm 0.5\%$ of full scale of selected range at 25 °C

TDS

Ranges:4 ranges, with automatic range selection and linearising software.

k=0.1 Sensor 1.000 ppM to 1000 ppM

k=1.0 Sensor 10.00 ppM to 10.00 ppK

k=10 Sensor 100.0 ppM to 100.0 ppK

Resolution.....0.1% of selected range

Accuracy..... $\pm 0.5\%$ of full scale of selected range at 25 °C

pH

Range0 to 14.00 pH

Resolution.....0.01 pH

Accuracy..... ± 0.01 pH

Temperature

Range-10.0 to 120.0 °C (Sensor limit 60 °C)

Resolution.....0.1 °C

Accuracy..... ± 0.2 °C

General Specifications

Temperature Compensation.....	Automatic, 0 to 100 °C Manual, 0 to 99.0 °C (pH only)
Cond. Sensor Span Range.....	k=0.1 Sensor : k=0.075 to k=0.133 k=1.0 Sensor : k=0.75 to k=1.33 k=10 Sensor: k=7.5 to k=13.3
pH Asymmetry Range	-1.00 to 1.00 pH
pH Slope Range	85.0 to 105%
Temp. Sensor Offset Range.....	-10.0°C to +10.0°C
Auto Standard Recognition	Conductivity 150 μS/cm, 1413 μS/cm, 2.76 mS/cm, 12.88 mS/cm, 58.0 mS/cm TDS 69.5 ppM, 2.00 ppK, 8.00 ppK, 36.0 ppK pH 4.00, 6.88, 7.00, 9.23, 10.00
Memory.....	2400 readings including date and time
Automatic Logging.....	User-set for one reading every 1 to 90 seconds, 1 to 90 minutes or 1 to 24 hours
RS232 Output.....	300, 1200, 9600 & 19200 baud. 8 bits, no parity, 1 stop bit, XON/XOFF Protocol.
Clock	Calendar clock displays date, month, hours, minutes & seconds. Year is recorded in memory and transmitted to the RS232 port, but is not displayed.
Battery Saver.....	On : Auto switch-off after 5 minutes Off : Continuous use Bar Graph display of battery charge level. Readout of battery voltage available for troubleshooting.

Good Laboratory Practices Date, Time and Value of last Conductivity, TDS, pH and Temperature calibration are stored, and can be recalled or sent to the RS232 port at any time.

Power 6V NiMH Rechargeable Battery for approx 40 hours operation.


Dimensions 187 x 110 x 51 mm

Mass Instrument only : Approx 440g
Full Kit : Approx 1.7kg

Environment Temperature : 0 to 45 °C
Humidity : 0 to 90 % R.H.

2. WP-81 Menu Structure

A detailed breakdown of the menu system of the **WP-81** is shown below. This diagram provides a quick reference for the menu functions available for the **WP-81**.



3. Operating Modes

3.1 **Selecting Conductivity or TDS Mode**

To select a Conductivity or TDS mode...

1. Select the Mode menu (Menu) → **F3:Mode**...

F1:Cond.	F2:pH
F3:TDS	F4:mV

2. Press (F1) to select Conductivity mode.
Press (F3) to select TDS mode.
Press (Menu) to quit and retain the current selection.

3.2 **Selecting pH or Millivolts (mV) Mode**

To select a pH or mV mode...

1. Select the Mode menu (Menu) → **F3:Mode**...

F1:Cond.	F2:pH
F3:TDS	F4:mV

2. Press (F2) to select pH mode.
Press (F4) to select mV mode.
Press (Menu) to quit and retain the current selection.

4. Conductivity Calibration

4.1 Calibration Procedure

1. Plug the Conductivity sensor into the **Conductivity/TDS** socket.
If a k=0.1 or k=10 sensor is being used, ensure that the **WP-81** is set to the correct k factor before using the instrument (see section 17).
2. Switch the meter on.
3. Select Conductivity Mode. (Menu) → **F3:Mode** → **F1:Cond.**)
4. Rinse the Conductivity electrode in distilled water. Shake off as much water as possible. Blot the outside of the electrode dry. **DO NOT BLOT THE ELECTRODE WIRES.**

5. Zero Calibration

Let the electrode dry in air.

Select Conductivity Calibration. (Menu) → **F1:Cal.** → **F1:Cond.**)

6. When the reading has stabilised at or near zero, press the (F1) key to calibrate.

A “ * “ will not be removed after a zero calibration.

7. Standard Calibration

Allowable Conductivity standards are 150uS/cm, 1413uS/cm, 2.76mS/cm, 12.88mS/cm and 58.0mS/cm, and should be selected according to your range of interest.


If the **WP-81** does not recognise the standard, it will display the message, “**NOT STD**” during calibration. Calibration will fail if this message is displayed.

For plastic bodied k=1 and k=10 sensors, place the electrode into a sample of Conductivity standard, so that it is immersed at least to the vent hole in the white plastic cover. The white plastic cover **MUST** be in place for correct readings.

For plastic bodied k=0.1 sensors, the white plastic cover **MUST** be removed for correct readings.

For glass bodied sensors, immerse the sensor at least to the vent hole in the glass body.

Refer to the diagrams over the page.



DO NOT place the electrode directly into the bottle of standard. Discard the used sample of standard after use. It is advisable to use a narrow sample vessel to minimise the use of standard solution.

8. Select Conductivity Calibration. (Menu) → **F1:Cal.** → **F1:Cond.**)
9. When the reading has stabilised, press the (F) key to calibrate.
The * will now be replaced by a decimal point, if calibration was successful.
10. The **WP-81** is now calibrated for Conductivity and is ready for use in this mode.

4.2 Calibration Notes

1. A Zero calibration should be performed at least monthly. In low conductivity applications (where a zero error is particularly significant) a zero calibration may have to be done weekly.
2. A Standard calibration should be performed at least weekly. Of course, more frequent calibration will result in greater confidence in results.
3. Conductivity and TDS calibration data is stored separately in memory. Ensure that the **WP-81** has been correctly calibrated for the mode in which it will be used. The **WP-81** does not require re-calibration when alternating between Conductivity and TDS modes, providing the instrument has been correctly calibrated for both.
4. All calibration information is retained in memory when the **WP-81** is switched off, even when the battery is removed. This information can be recalled or printed later using the GLP function (see section 9).
5. The **WP-81** displays the value of the standard to which it will attempt to calibrate. Ensure that the standard value displayed corresponds to the standard that you are using.
6. If the **WP-81** does not recognise the standard, it will display the message, "**NOT STD**" during calibration. Calibration will fail if this message is displayed.

4.3 Calibration Messages

1. If a Zero calibration has been successfully performed, the **WP-81** will display the following message, and the zero value of the electrode. For example...

```
Calibrate OK  
Zero= 0.00 uS
```

2. If a Standard calibration has been successfully performed, the **WP-81** will display the following message, and the k factor of the electrode. For example...

```
Calibrate OK  
k= 1.00
```

3. If a Standard calibration has failed, the **WP-81** will display the following message, and the failed k factor of the electrode. For example..

```
Calibrate Fail  
k= 1.50
```

5. TDS Calibration

5.1 Calibration Procedure

1. Plug the TDS sensor into the **Conductivity/TDS** socket.
If a k=0.1 or k=10 sensor is being used, ensure that the **WP-81** is set to the correct k factor before using the instrument (see section 17).

2. Switch the meter on.

3. Select TDS Mode. (Menu) → **F3:Mode** → **F3:TDS**)

4. Rinse the TDS electrode in distilled water. Shake off as much water as possible. Blot the outside of the electrode dry. **DO NOT BLOT THE ELECTRODE WIRES.**

5. Zero Calibration

Let the electrode dry in air.

Select TDS Calibration. (Menu) → **F1:Cal.** → **F1:TDS**)

6. When the reading has stabilised at or near zero, press the (F1) key to calibrate.

The * will not be removed after a zero calibration.

7. Standard Calibration

Allowable standards are 69.5ppM, 2.00ppK, 8.00ppK, and 36.0ppK, and should be selected according to your range of interest.


If the **WP-81** does not recognise the standard, it will display the message, "**NOT STD**" during calibration. Calibration will fail if this message is displayed.

For plastic bodied k=1 and k=10 sensors, place the electrode into a sample of TDS standard, so that it is immersed at least to the vent hole in the white plastic cover. The white plastic cover **MUST** be in place for correct readings.

For plastic bodied k=0.1 sensors, the white plastic cover **MUST** be removed for correct readings.

For glass bodied sensors, immerse the sensor at least to the vent hole in the glass body.

Refer to the diagrams over the page.



DO NOT place the electrode directly into the bottle of standard. Discard the used sample of standard after use. It is advisable to use a narrow sample vessel to minimise the use of standard solution.

8. Select TDS Calibration. (Menu) → **F1:Cal.** → **F1:TDS**)
9. When the reading has stabilised, press the (F1) key to calibrate.
10. The * will now be replaced by a decimal point, if calibration was successful.
11. The **WP-81** is now calibrated for TDS and is ready for use in this mode.

5.2 Calibration Notes

1. A Zero calibration should be performed at least monthly. In low TDS applications (where a zero error is particularly significant) a zero calibration may have to be done weekly.
2. A Standard calibration should be performed at least weekly. Of course, more frequent calibration will result in greater confidence in results.
3. TDS and Conductivity calibration data is stored separately in memory. Ensure that the **WP-81** has been correctly calibrated for the mode in which it will be used. The **WP-81** does not require re-calibration when alternating between TDS and Conductivity modes, providing the instrument has been correctly calibrated for both.
4. All calibration information is retained in memory when the **WP-81** is switched off, even when the battery is removed. This information can be recalled or printed later using the GLP function (see section 9).
5. The **WP-81** displays the value of the standard to which it will attempt to calibrate. Ensure that the standard value displayed corresponds to the standard that you are using.
6. If the **WP-81** does not recognise the standard, it will display the message, "**NOT STD**" during calibration. Calibration will fail if this message is displayed.

5.3 Calibration Messages

1. If a Zero calibration has been successfully performed, the **WP-81** will display the following message, and the zero value of the electrode. For example...

```
Calibrate OK  
Zero= 0.00 ppm
```

2. If a Standard calibration has been successfully performed, the **WP-81** will display the following message, and the k factor of the electrode. For example...

```
Calibrate OK  
k= 1.00
```


3. If a Standard calibration has failed, the **WP-81** will display the following message, and the failed k factor of the electrode. For example..

```
Calibrate Fail  
k= 1.50
```

6. pH Calibration

6.1 Calibration Procedure

1. Plug the pH sensor into the **pH** socket. Temperature measurements are made via the Conductivity sensor, so this needs to be connected for Automatic Temperature Compensation.
2. Switch the meter on.
3. Ensure that temperature has been calibrated or manually set (see sections 8.1 & 8.4). If the decimal point in the temperature reading is shown by a “ * “, then the temperature readout is not calibrated.
4. Remove the wetting cap from the pH sensor. Rinse the pH and Conductivity sensors in distilled water and blot them dry.
5. Ensure that you are using buffers which have been selected for automatic buffer recognition. See section 18 for a detailed explanation.
6. Place both electrodes into a small sample of pH6.88 (or pH7.00) buffer, so that the bulb and reference junction are both covered, as per the diagram below. **DO NOT** place the electrodes directly into the buffer bottle.



7. Select pH Calibration. (Menu → **F1:Cal.** → **F2:pH**)
8. When the reading has stabilised, press the (F1) key to calibrate. If a 1 point calibration has been performed, the * will not be removed until a full 2 point calibration has been performed.
9. Rinse the pH and Conductivity electrodes in distilled water and blot them dry.
10. Place both sensors into a small sample of pH4.00, pH9.23 or pH10.00 Buffer, so that the bulb and reference junction are both covered, as per the diagram in step 6, above. **DO NOT** place the electrodes directly into the buffer bottle.

NOTE: pH9.23 and pH10.00 buffers are highly unstable. Avoid using these buffers if possible. Discard immediately after use.

11. Select pH Calibration. (Menu → **F1:Cal.** → **F2:pH**)
12. When the reading has stabilised, press the (F1) key to calibrate. The * will now be replaced by a decimal point, if calibration was successful.
13. The **WP-81** is calibrated for pH and is ready for taking pH measurements. Discard the used samples of buffer.

6.2 Calibration Notes

1. A 1-point calibration should be performed at least weekly. In applications where the electrode junction can become blocked, such as dairy products, mining slurries etc., a 1-point calibration may have to be done daily.
2. A full 2-point calibration should be performed at least monthly. Of course, more frequent calibration will result in greater confidence in results.
3. All calibration information is retained in memory when the **WP-81** is switched off, even when the battery is removed. This information can be recalled or printed later using the GLP function (see section 9).
4. The **WP-81** displays the value of the pH buffer to which it will attempt to calibrate. Ensure that the buffer value displayed corresponds to the buffer that you are using.

6.3 Calibration Messages

1. If a 1-point calibration has been successfully performed, the **WP-81** will display the following message, and the asymmetry of the electrode. For example...

```
1 Point Cal. OK
Asy= 0.10pH
```

2. If a 1-point calibration has failed, the **WP-81** will display the following message, and the failed asymmetry value of the electrode. For example...

```
1 Point Cal. Fail
Asy= 1.50pH Hi
```

or :

```
1 Point Cal. Fail
Asy= -1.50pH Lo
```

3. If a 2-point calibration has been successfully performed, the **WP-81** will display the following message, and the asymmetry and slope of the electrode. For example...

```
2 Point Cal. OK
Asy= 0.10pH
```

then :

```
2 Point Cal. OK
Slope=100.0%
```

4. If a 2-point calibration has failed, the **WP-81** will display the following message, and the failed slope value of the electrode. For example...

```
2 Point Cal. Fail
Slope=130.0% Hi
```

or :

```
2 Point Cal. Fail
Slope= 70.0% Lo
```


7. mV Calibration

The mV section is factory calibrated. There is no user-calibration facility for this mode.

8. Temperature Calibration

8.1 Calibration Procedure

1. Plug the Conductivity/Temperature sensor into the **Conductivity / TDS** socket. A separate temperature sensor can also be used in place of the Conductivity sensor for temperature readout.
2. Switch the meter on.
3. Place the sensor into a beaker of room temperature water, alongside a good quality mercury thermometer. Stir the sensor and the thermometer gently to ensure an even temperature throughout the beaker.
4. Select Temperature Calibration. (Menu) → **F1:Cal.** → **F3:Temp**)
5. The reading from the probe is now displayed on the right of the display, and the value you are going to set is shown on the left. For example...



The screenshot shows a green display with two lines of text. The top line reads "→ 25.0← 20*0°c" and the bottom line reads "↑↓:Set F1:Cal.". The numbers 25.0 and 20*0°c are on the right side, and the text "↑↓:Set" and "F1:Cal." are on the left side.

6. When the reading on the right has stabilised, press the ▲ and ▼ keys until the reading on the left shows the same temperature as the mercury thermometer.
7. Press the F1 key to calibrate the temperature readout.
The * will now be replaced by a decimal point, if calibration was successful.
Alternatively, press the Menu key to abort temperature calibration.

8.2 Calibration Notes

1. Temperature calibration information is retained in memory when the **WP-81** is switched off, even when the battery is removed. This information can be recalled or printed later using the GLP function (see section 9).
2. Temperature does not need to be re-calibrated unless the sensor is replaced or the meter is initialised.

8.3 Calibration Messages

1. If a temperature calibration has been successfully performed, the **WP-81** will display the following message and the offset value of the probe. For example...

```
Calibrate OK
Offset= 1.0°C
```


2. If a temperature calibration has failed, the **WP-81** will display the following message, and the failed offset value of the probe.

```
Calibrate Fail
Offset= 10.5°C
```



8.4 Manual Temperature Setting


If the Conductivity/TDS/Temperature sensor is not connected, and a temperature sensor is not used in its place, the temperature of the sample solution must be set manually for accurate pH measurements. A separate thermometer will be required for this.

NOTE: The Conductivity sensor has a separate sensor built in for automatic temperature compensation for Conductivity and TDS.

1. Switch the meter on.
2. Measure the temperature of the sample.
3. Select Temperature Calibration. ( → **F1:Cal.** → **F3:Temp**)
4. The current temperature setting is now displayed.

```
→ 25.0← Man Temp
↑↓:Set   F1:Save
```

5. Press the  and  keys until the display shows the temperature of the sample.
6. Press the **F1** key to save the temperature value.


Alternatively, press the  key to quit and retain the current setting.

9. Good Laboratory Practices (GLP)




The **WP-81** keeps a record of the date and time of the last Conductivity, TDS, pH and Temperature calibrations as part of GLP guidelines. The zero and span values for Conductivity and TDS are stored separately.

9.1 To recall GLP information on the display

1. Switch the meter on.
2. Select the GLP menu. (Menu) → **F4:Setup** → **F3:GLP**)
3. Select **F1:Recall** from the menu.
4. The instrument model, firmware version number, and instrument serial number are displayed, along with a prompt describing how to scroll through the GLP information.



```
WP81 V5.1 S1234
↓:More ↑:Back
```

5. Press the  key to sequentially scroll through the GLP information for all parameters. Press the  key to scroll back to previous data. The sequence of information displayed is shown below. Press  to abort at any time.

```
WP81 V5.1 S1234
↓:More ↑:Back
```

:↓ :↑

```
Cond Zero 0.00uS
@ 31/12/03 11:00
```

:↓ :↑

```
k=1.00 @ 2.76mS
@ 31/12/03 11:10
```

:↓ :↑

```
TDS Zero 0.00ppM
@ 31/12/03 11:20
```

:↓ :↑



```
k=1.00 @ 2.00ppK
@ 31/12/03 11:30
```

:↓ :↑



```
pH Asy 0.10pH
@ 31/12/03 11:40
```

:↓ :↑

```
pH Slope 100.0%
@ 31/12/03 11:50
```

:↓ :↑

```
Temp Offset 0.1°C
@ 31/12/03 12:00
```

:↓ :↑

Exit

9.2 Failed Calibration


If calibration has failed, the GLP function will reset the date and time to zero. The **WP-81** still shows the results of the last successful calibration. For example...

Cond Zero 0.00uS @ 00/00/00 00:00	k=1.00 @ 00/00/00 00:00
Asy 0.10pH @ 00/00/00 00:00	Slope 100.0% @ 00/00/00 00:00
Temp Offset 1.0°C @ 00/00/00 00:00	

Note that these calibration values are still used if further measurements are taken without re-calibrating.

9.3 Printing GLP Information to the RS232 Port

The GLP information stored in the instrument's memory can be sent to a printer or PC via the RS232 port. Switch the meter on.


1. Ensure that the **WP-81** RS232 cable is connected to the instrument and to the printer or PC.
2. Send the GLP information to the RS232 port.
() → **F4:Setup** → **F3:GLP** → **F3:Print**)
3. The GLP information is sent to the RS232 port in formatted ASCII text. For example...

```
WP81 V5.1 S1234 @ 31/12/03 12:00
Conductivity Zero= 0.00uS @ 31/12/03 11:00
Conductivity k= 1.00 @ 2.76mS @ 31/12/03 11:10
TDS Zero= 0.00ppM @ 31/12/03 11:20
TDS k= 1.00 @ 36.0ppK @ 31/12/03 11:30
pH Asy= 0.00pH @ 31/12/03 11:40
pH Slope= 100.0% @ 31/12/03 11:50
Temperature Offset= 1.0oC @ 31/12/03 12:00
ENDS
```

9.4 Instrument Serial Number

In case the serial number that is fitted to the rear of the **WP-81** is removed or becomes illegible, it is also available on the **WP-81** display.

- The serial number is displayed at turn-on, for example...



The image shows a green rectangular display area with black text. The top line reads 'WP81 V5.1 S1234' and the bottom line reads 'Cond TDS pH Temp'. The text is centered and appears to be a status or information screen from the device.

where **S1234** is the serial number.

- The serial number is display when recalling the GLP information (section 9.1).
- The serial number is included on the print-out of GLP information (section 9.3).

9.5 Additional GLP Features

Another GLP requirement is to record the date and time of every reading. The **WP-81** does this for you when readings are recorded either with the Notepad function (section 10) or the Automatic Logging function (section 11).

10. Notepad Function

10.1 Recording Readings into the Notepad

To record readings into the Notepad memory:

1. Press **F1** in normal display mode. The display should now look like this...

2.76mS	7.00pH
F1: 1	12:00:00

or :

2.00ppK	7.00pH
F1: 1	12:00:00

2. If you now press **F1**, the Conductivity/TDS, pH, Temperature, Date and Time will be recorded into the notepad, and labelled as reading number 1.
3. Repeat steps 1 & 2 as often as required. The maximum number of readings that can be stored in the Notepad is 2400.

10.2 Recalling Records from the Notepad

To recall records from the Notepad onto the **WP-81** display:

1. Select the Notepad menu (**Menu** → **F2:Notepad**)
2. Select **F1:Recall** from the menu.
3. Record number 1 is now displayed, for example...


2.76mS	7.00pH
25.0°C	1 F2:Clk

4. Press **F2** to alternatively display the date and time or the data for this record.
Press **▲** to move forward through the records.
Press **▼** to move backward through the records.
Press and hold the **▲** or **▼** keys to roll rapidly through the readings.

10.3 Erasing Records from the Notepad

To erase all records from the Notepad:

1. Select the Notepad menu (Menu → **F2:Notepad**)
2. Select **F2:Erase** from the menu.
3. The **WP-81** now asks if you are sure that you wish to erase all records...



```
Erase, You Sure?
F1:Yes  F2:No
```

4. Press **F1** to erase all records from the Notepad
Press **F2** to quit without erasing the records from the Notepad.

10.4 Printing Records from the Notepad to the RS232 Port

1. Connect one end of the RS232 cable to the **Charger/RS232** socket of the **WP-81**. The charger, optional solar panel, or optional battery leads can be connected into the spare socket on the cable for long term use, if required.
2. Connect the other end of the RS232 cable to an RS232 Printer, or to COM1 or COM2 of a PC.
3. Ensure that the baud rate for the printer or PC and the **WP-81** are the same.

If necessary, alter the baud rate of the **WP-81** (see section 12.1).


The **WP-81** uses XON/XOFF protocol. Ensure that the printer or PC is set accordingly.

4. Select the Notepad menu. (Menu → **F2:Notepad**)
5. Select **F3:Print** from the menu.
Printing starts as soon as **F3** is pressed. The display shows the word "**Printing**" until printing is completed.



11. Automatic Datalogging

The **WP-81** can automatically log records into the Notepad. First the logging period must be programmed, then automatic logging can be started and stopped as required. The clock must be set before attempting Automatic Datalogging.

1. Select the Program menu. (Menu) → **F2:Notepad** → **F4:Prog.**)
2. The display should now look like this...




The screenshot shows a green rectangular display with black text. The text is arranged in two lines. The first line reads "→00← F1:Min F2:Sec". The second line reads "↑↓:Period F3:Hr".

3. Use the  and  keys to set the period at which the **WP-81** will automatically log records.
4. When the logging period has been correctly set, select whether this period is in minutes, seconds or hours.
Press **F1** to save the period as minutes.
Press **F2** to save the period as seconds.
Press **F3** to save the period as hours.
For example, if the period was set to **05**, followed by **F2**, then the **WP-81** will automatically log a record every 5 seconds.
5. The **WP-81** will ask if the records are to be logged into the Notepad, or sent directly to the RS232 port.
Press **F1** to log records into the Notepad (maximum of 2400 readings).
Press **F3** to send records directly to the RS232 port.
6. The automatic logging function is now programmed, and can be started and stopped as required.

Continued over the page...

7. To start automatic logging, press F3 in normal display mode.


If the **WP-81** is logging into the Notepad, the display will look like this...



```
2.76mS    7.00pH
Log#  1 12:00:00
```

The log number will increment and the **WP-81** will beep each time a reading is recorded.

If the **WP-81** is sending records directly to the RS232 port, the display will look like this...



```
2.76mS    7.00pH
Sending 12:00:00
```

The **WP-81** will beep each time a record is sent to the RS232 port.

8. Press F3 to stop automatic logging.


Notes:

1. The clock must be set before the **WP-81** will allow automatic logging to start. The message "**Clock Not Set**" is displayed if the clock is not set.
2. The Battery Saver function (section 14) is disabled while the meter is in Automatic Datalogging mode, to stop the meter switching off while logging data. Even when the memory is full and the meter stops logging, the Battery Saver function is still disabled. This allows the data to be downloaded and the memory to be reset remotely.

12. RS232 Port

12.1 Setting the Baud Rate

1. Select the RS232 Set-up menu (Menu) → **F4:Setup** → **F4:Baud**)
2. The available baud rates are listed on the display...



```
F1:300  F2:1200
>F3:9600 F4:19200
```

The arrow shows the current selection.

3. Press (F1) to select 300 baud
Press (F2) to select 1200 baud
Press (F3) to select 9600 baud.
Press (F4) to select 19200 baud.
Press (Menu) to quit and retain the current setting.

12.2 Sending Readings to the RS232 Port

Press (F3) to instantly send readings to the RS232 port whenever the **WP-81** is in normal display mode. This function is disabled if the automatic logging period is set to greater than zero (see section 11).

Records can be sent directly to the RS232 port rather than stored in memory during automatic datalogging. See section 11 for details.

12.3 RS232 Configuration

The **WP-81** RS232 configuration is 8 bits, No Parity, 1 Stop Bit, XON/XOFF Protocol.

12.4 Communication and Statistical Software

Communication between the **WP-81** and a PC can be handled with any RS232 communication software. **WinTPS** RS232 communication software for Windows® 95 and later is optionally available (part number 130086).

Once the data is saved to disk, the next problem is how to use it. The data sent by the **WP-81** is formatted in fixed-width columns that can be imported by programs such as Microsoft® Excel® and Lotus 123®.

Information on how to use the software and import data is provided in the manual provided with the **WinTPS** CD-ROM.

12.5 Commands

The following commands can be sent from a PC to the **WP-81**. Note that <cr> denotes carriage return and <lf> denotes a line feed.

Action	Command	Notes
Request current data	?D<cr>	Returns the current Conductivity/TDS, pH, Temperature, date and time from the WP-81 . The log number returned is set to Zero.
Request logged data	?R<cr>	Returns all logged records from the WP-81 memory. The data ends with the message ENDS <cr>
Erase logged data	?E<cr>	Erases all logged records from the WP-81 memory. Returns the message ERASED <cr> to confirm that the records have been erased.
Request status information	?S<cr>	Returns the model name, firmware version number, instrument serial number and number of logged readings in memory, eg: WP81♦♦V1.0♦S1234♦9999 <cr>, where ♦ are spaces. Note that the number of logged readings is right-justified.
Request GLP information	?G<cr>	Returns all calibration GLP information, plus the instrument model and current date (see section 12.6 for data format and handshaking).

12.6 Data Format

Data is returned to the RS232 Port by the **WP-81** in the following format. Please note that a “ ♦ ” shown anywhere in this section denotes one space.

LLLL♦CCCCCCccc♦PPPPPPppp♦TTTTTTttt♦dd/mm/yy♦hh:mm:ss

where...

LLLL	is the Log Number. Maximum 4 characters, right justified. The WP-81 sends Zero for instant readings (section 12.2)
CCCCCC	is the Conductivity or TDS Data. Maximum 6 characters, right justified.
ccc	is the unit description, either “ uS♦ ”, “ mS♦ ”, “ ppM ” or “ ppK ”.
PPPPPP	is the pH or mV Data. Maximum 6 characters, right justified.
ppp	is the unit description, sent as “ pH♦ ” or “ mV♦ ”
TTTTTT	is the Temperature Data. Maximum 6 characters, right justified.
ttt	is the Temperature unit description. The WP-81 sends “ oC♦ ” for real temperature data, or “ oCm ” when manual temperature compensation is being used.
dd/mm/yy	is the date, month and year data.
hh:mm:ss	is the hours, minutes and seconds data.

Notes

1. When requested by a PC with the ?D or ?R commands (section 12.5), the data is terminated with a carriage return.
2. When the data is sent by the **WP-81** using the Print function (section 10.4) or the Instant Send function (section 12.2), the data ends with a carriage return and a line feed.

12.7 GLP Data Format

GLP information is returned as 8 lines terminated by a carriage return. When using the "?G" command (section 12.5), the computer must respond with a character after receiving each line. For example...

```
WP81 V5.1 S1234 @ 31/12/03 12:00
Conductivity Zero= 0.00uS @ 31/12/03 11:05
Conductivity k= 1.00 @ 2.76mS @ 31/12/03 11:10
TDS Zero= 0.00ppM @ 31/12/03 11:15
TDS k= 1.00 @ 36.0ppK @ 31/12/03 11:20
pH Asy= 0.00pH @ 31/12/03 11:25
pH Slope= 100.0% @ 31/12/03 11:30
Temperature Offset= 1.0oC @ 21/12/03 11:35
ENDS
```


13. Importing Data into Microsoft Excel

The following procedure details the method for importing a **WP-81** text data file into Microsoft® Excel®.

1. Start Microsoft® Excel® and select File → Open
2. In the “Files of type:” pull-down box, choose “Text Files (*.prn; *.txt; *.csv)”.
3. Navigate to the folder where your data file is stored and double-click it to start the Text Import Wizard.


Note: The default data folder for the WinTPS software is “C:\My Documents\WinTPS”.

4. In step 1 of the Text Import Wizard select “Fixed width”, as per the sample screen below, then press “Next >”.



Continued over the page...

5. Step 2 of the Text Import Wizard allows you to select the points at which each data field will break into a new column. The sample screens below show where TPS recommends the breaks be inserted. Press "Next >" after the column breaks have been inserted.



6. Simply press "Finish" at step 3 of the Text Import Wizard. TPS recommends that the data format for each column be set once the data is in spreadsheet format.

For help on formatting the data columns, charting, graphing or other operations please consult the Microsoft® Excel® help file. Alternatively please contact TPS and we will try to provide further assistance.

14. Battery Saver Function

The **WP-81** is equipped with a battery saver function. If no button has been pressed for five minutes, the unit beeps and flashes the display for 20 seconds, and then shuts off. This function can be switched off for continuous use.

To enable or disable the battery saver function:

1. Switch the meter on.
2. Select Battery Saver Set-up (Menu) → **F4:Setup** → **F1:Batt**)
3. The battery saver menu is now displayed. For example...




The arrow indicates the current selection.


The bar graph and percentage indicate the approximate level of charge in the battery.

4. Press (F1) to disable the battery saver function for continuous use.
Press (F2) to enable the battery saver function. The meter will switch itself off if no key has been pressed for five minutes.
Press (Menu) to quit the battery saver menu and retain the current setting.

Notes:


1. For troubleshooting purposes, the battery volts can also be displayed in the battery saver menu. Press (F3) to display battery volts.
2. The  symbol flashes when the battery volts drops below 5.60 volts. At 5.00 volts the meter turns itself off.
3. The Battery Saver function is disabled while the meter is in Automatic Datalogging mode (section 11), to stop the meter switching off while logging data. Even when the memory is full and the meter stops logging, the Battery Saver function is still disabled. This allows the data to be downloaded and the memory reset remotely.

15. Recharging the Battery

The  symbol flashes when the battery drops below 5.60 volts. The battery should be recharged at this point. If the battery is not recharged, the **WP-81** will switch itself off when the battery drops below 5.00 volts.

To recharge the battery...


1. Plug the battery charger, solar panel, or car cigarette lighter adaptor into the **Charger/RS232** socket. **DO NOT** plug into the **Conductivity / TDS** socket, as this will damage the **WP-81**.
2. Charge for approximately 8 hours for full capacity. The **WP-81** has special circuitry to prevent overcharging, so the charger can be used continuously.

To ensure optimum battery life and capacity, the **WP-81** should only be charged once the  symbol starts to flash.

16. Clock Function

16.1 Setting the Clock

1. Select the Clock Set-up menu (Menu) → **F4:Setup** → **F2:Clock**
2. The display now shows the current date and time. The cursor starts at the day.



31/12/03 12:00
F1:< F2:> ↑↓:Set

3. Press the (▲) and (▼) keys until the day is correct.
4. Press (F2) to move to the month. Press the (▲) and (▼) keys until the month is correct.
5. Press (F2) to move to the year. Press the (▲) and (▼) keys until the year is correct.
6. Press (F2) to move to the hour. Press the (▲) and (▼) keys until the hour is correct.
7. Press (F2) to move the cursor to the minutes. Press the (▲) and (▼) keys until the minutes are correct.
8. Check that the date and time are correct.

Press (F2) to save the settings.

If any changes are needed, press the (F1) key to move left to the desired position.

Press (Menu) to quit without resetting the clock.

Notes

1. The **WP-81** does not test for a valid day of the month when setting the clock (eg: attempting to enter 31/02/03 is not corrected).
2. The **WP-81** does test for leap years.

16.2 Displaying or Hiding the Clock

The time is normally displayed along with the Conductivity/TDS, pH and Temperature readings.

Press (F2) in normal display mode to hide the time.

Press (F2) again to display the time plus the date.

The temperature reading replaces the date after 5 seconds.

17. Selecting k=0.1 or k=10 Sensors

The **WP-81** automatically recognises a k=1.0 sensor. The **WP-81** **does not** automatically recognise k=0.1 or k=10 sensors. When a k=0.1 or k=10 sensor is used, the **WP-81** must be set to the correct k factor before use. The following procedure describes how to select a k=0.1 or k=10 sensor.

1. Switch the meter **OFF**.
2. Connect the k=0.1 or k=10 sensor.
3. Press **and HOLD** the **F2** key while switching the meter back on.
4. The k factor selection menu is now displayed (only if the k=0.1 or k=10 sensor is connected)...

```
Select      F1:k=0.1
k factor >F2:k=10
```

The arrow indicates the current selection.

5. Press **F1** to select a k=0.1 sensor.
Press **F2** to select a k=10 sensor.
Press **Menu** to quit buffer selection and retain the current setting.


Notes

1. The manual k factor selection is kept in memory when the meter is switched off, even if the battery is removed.
2. The manual k factor selection is reset to k=10 during initialisation.
3. The **WP-81** will always automatically recognise a k=1.0 sensor, regardless of the manual k factor selection.
4. Calibration settings for k=0.1, k=1.0 and k=10 sensors are **NOT** stored separately. The **WP-81** requires re-calibration when a new k factor sensor is connected.

18. Selecting Buffers for Auto Buffer Recognition

The **WP-81** is factory set to automatically recognise pH4.00, pH6.88 and pH9.23 buffers. However, some users may prefer to use pH7.00 instead of pH6.88 and pH10.00 instead of pH9.23. The following procedure describes how to set which of these buffers are automatically recognised at calibration.


1. Switch the meter **OFF**.
2. Press **and HOLD** the F1 key while switching the meter back on.
3. The buffer selection menu is now displayed.



```
Select >F1:6.88pH
Buffer F2:7.00pH
```

The arrow indicates the current selection.

4. Press F1 to select pH6.88 as the primary buffer.
Press F2 to select pH7.00 as the primary buffer.
Press Menu to quit buffer selection and retain the current setting.
5. The display will now show the currently selected high pH buffer.



```
Select >F1:9.23pH
Buffer F2:10.0pH
```

The arrow indicates the current selection.


6. Press F1 to select pH9.23 as the high pH buffer.
Press F2 to select pH10.00 as the high pH buffer (the display shows 10.0 for the latter, but this buffer is stored as pH10.00).
Press Menu to quit buffer selection and retain the current setting.
7. The setting is kept in memory when the meter is switched off, even if the battery is removed. The buffers are re-set to pH6.88 and pH9.23 during initialisation.

Note: pH6.88 buffer is a DIN 19266 and NBS Primary-standard pH solution. Its use is highly recommended for the most accurate possible results. If pH7.00 buffer is used, ensure that it is manufactured to 0.01pH accuracy. pH7.00 buffer has a buffer capacity less than half that of pH6.88 buffer and is therefore much less stable.

19. Initialising the WP-81

If the calibration settings of the **WP-81** exceed the allowable limits, the unit may need to be initialised to factory default values. This action may be required if the electrode is replaced.

To initialise the **WP-81**:

1. Switch the **WP-81** OFF.
2. Press **and HOLD** the  key while switching the **WP-81** back on.
3. The following messages should be displayed...

```
Initialized
MUST ReCalibrate
```

then :

```
WP81s V5.1 S1234
Cond TDS pH Temp
```

(The "s" after **WP-81** is shown when the RS232 serial port option is fitted)

4. The meter then displays Conductivity, pH and Temperature. Note that the decimal points have been replaced with a " * ", to indicate that the unit requires re-calibration.

Notes:


1. When the **WP-81** is initialised, the manual k factor selection is re-set to k=10. See section 17 if you wish to select a k=0.1 sensor.
2. When the **WP-81** is initialised, the automatically recognised buffers are re-set to pH4.00, pH6.88 and pH9.23. See section 18 if you prefer to use pH7.00 instead of pH6.88 and/or pH10.00 instead of pH9.23.

20. Instrument firmware version number.

If you need to phone or fax TPS for any further technical assistance, the version number of your **WP-81** firmware may of benefit to us. The version number is displayed by the **WP-81** at turn-on.

21. Troubleshooting

21.1 General Errors

Error Message	Possible Causes	Remedy
Factory Cal. Failed See Handbook	The EEPROM chip which contains the factory calibration information has failed.	The unit must be returned to TPS for service.
Memory Failed Calibration Lost Initialised MUST ReCalibrate	User calibration settings have been lost or corrupted.	Re-calibrate the instrument. A full 2-point calibration will be required for Conductivity, TDS & pH (sections 4.1, 5.1 & 6.1) and a 1 point calibration for temperature (section 8.1).
Flashing  symbol.	Battery is below 5.60 volts.	Recharge the battery. Note that the unit will switch itself off when the battery falls below 5.00 volts.
Meter displays the word OFF , and switches off.	Battery is below 5.00 volts.	Recharge the battery. If this fails, check the charger. If charger OK, replace the battery.
Meter will not turn on.	<ol style="list-style-type: none"> 1. Battery is exhausted. 2. Faulty Instrument 	Recharge the battery. If this fails, check the charger. If charger OK, replace the battery. Return to factory for repair.
Battery does not charge up when charger is connected.	<ol style="list-style-type: none"> 1. Faulty battery charger or faulty battery. 2. Faulty instrument. 	Connect the charger and switch the power on. Display the battery volts in the battery saver menu (section 14). If the battery volts are increasing then the charger is OK. If the battery volts do not increase, then the charger is faulty. Replace the charger or the battery, as required. Return to factory for repair.

21.2 Conductivity and TDS Troubleshooting

Symptom	Possible Causes	Remedy
Unit fails to calibrate, even with new electrode.	Calibration settings outside of allowable limits due to previous failed calibration.	Initialise the unit. See section 19.
Unit attempts Span calibration instead of Zero calibration.	Electrode has Zero error.	Thoroughly rinse electrode in distilled water and allow to completely dry in air before attempting zero calibration. If instrument does not calibrate at Zero with electrode disconnected, then the instrument is faulty.
Standard calibration fails, and k factor is greater than 0.133, 1.33 or 13.3, (depending on k factor of sensor).	<ol style="list-style-type: none"> 1. Electrode is not immersed deeply enough. 2. Electrode may have a build-up of dirt or oily material on electrode wires. 3. Platinum-black coating has worn off. 4. Standard solution is inaccurate. 5. Electrode is faulty. 6. Faulty instrument. 7. k-factor incorrectly set if using k=0.1 or k=10 sensor. 	<p>Immerse electrode at least to the vent hole in the white plastic cover.</p> <p>Clean electrode, as per the instructions detailed in section 22.1.</p> <p>Electrode requires replatinisation. Return to the factory, or see details in section 22.2.</p> <p>Replace standard solution.</p> <p>Return electrode to factory for repair or replacement.</p> <p>Return to factory for repair.</p> <p>Set the correct k-factor, as per section 17.</p>

Continued next page...

Conductivity and TDS Troubleshooting, continued...

Standard calibration fails, and k factor is less than 0.075, 0.75 or 7.5, (depending on k factor of sensor).	<ol style="list-style-type: none"> 1. White protective cover is not fitted or upside down. 2. Standard solution is inaccurate. 3. Electrode may have a build-up of conductive material, such as salt. 4. Electrode is faulty. 5. Faulty instrument. 6. k-factor incorrectly set if using k=0.1 or k=10 sensor. 	<p>The white protective cover MUST be fitted for correct readings. The vent hole must be towards the cable end of the electrode.</p> <p>Replace standard solution.</p> <p>Clean electrode, as per the instructions detailed in section 22.1.</p> <p>Return electrode to factory for repair or replacement.</p> <p>Return to factory for repair.</p> <p>Set the correct k-factor, as per section 17.</p>
Inaccurate readings, even when calibration is successful.	<ol style="list-style-type: none"> 1. Electrode may have a build-up of dirt or oily material on electrode wires. 2. Platinum-black coating has worn off. 	<p>Clean electrode, as per the instructions detailed in section 22.1.</p> <p>Electrode requires replatinisation. Return to the factory, or see details in section 22.2.</p>
Readings drift.	<ol style="list-style-type: none"> 1. Electrode may have a build-up of dirt or oily material on electrode wires. 	<p>Clean electrode, as per the instructions detailed in section 22.1.</p>
Readings are low or near zero.	<ol style="list-style-type: none"> 1. Electrode may have a build-up of dirt or oily material on electrode wires. 2. Electrode is not immersed deeply enough. 3. Electrode is faulty. 4. Faulty instrument. 5. k-factor incorrectly set if using k=0.1 or k=10 sensor. 	<p>Clean electrode, as per the instructions detailed in section 22.1.</p> <p>Immerse electrode at least to the vent hole in the white plastic cover.</p> <p>Return electrode to factory for repair or replacement.</p> <p>Return to factory for repair.</p> <p>Set the correct k-factor, as per section 17.</p>

21.3 pH Troubleshooting

Symptom	Possible Causes	Remedy
Unit fails to calibrate, even with new probe.	Calibration settings outside of allowable limits due to previous failed calibration.	Initialise the unit. See section 19.
1 Point calibration fails (Asymmetry is greater than +/- 1.00 pH).	<ol style="list-style-type: none"> Reference junction blocked. Reference electrolyte contaminated. 	<p>Clean reference junction, as per instructions supplied with the electrode.</p> <p>Flush with distilled water and replace electrolyte.</p>
2 Point calibration fails (Slope is less than 85.0%).	<ol style="list-style-type: none"> Incorrect primary buffer. Glass bulb not clean. Electrode is aged. Connector is damp. Buffers are inaccurate. 	<p>Ensure that you are using the primary pH buffer for which the WP-81 has been set (see section 18).</p> <p>Clean glass bulb as per instructions supplied with the electrode.</p> <p>Attempt rejuvenation, as per instructions supplied with the electrode. If not successful, replace electrode.</p> <p>Dry in a warm place.</p> <p>Replace buffers.</p>
Unstable readings.	<ol style="list-style-type: none"> Electrolyte chamber needs to be refilled. Reference junction blocked. Glass bulb not clean. Bubble in glass bulb. Faulty connection to meter. Reference junction not immersed. KCl crystals around reference junction, inside the electrolyte chamber. 	<p>Refill with saturated KCl filling solution.</p> <p>Clean reference junction, as per instructions supplied with the electrode.</p> <p>Clean glass bulb as per instructions supplied with the electrode.</p> <p>Flick the electrode to remove bubble.</p> <p>Check connectors. Replace if necessary.</p> <p>Ensure that the bulb AND the reference junction are fully immersed.</p> <p>Rinse electrolyte chamber with warm distilled water until dissolved. Replace electrolyte.</p>

Continued next page...

pH and mV Troubleshooting, continued...

Inaccurate readings, even when calibration is successful.	Reference junction blocked.	Clean reference junction, as per instructions supplied with the electrode.
Displays 7.00 for all solutions.	Electrical short in connector.	1. Check connector. Replace if necessary. 2. Replace electrode.
Displays 4-5 pH for all solutions.	Glass bulb or internal stem cracked.	Replace electrode.

21.4 Temperature Troubleshooting

Symptom	Possible Causes	Remedy
Displays "OVR°C" when electrode is plugged in.	1. Faulty electrode. 2. Faulty instrument.	Fit new electrode, part number 122201. Return to factory for repair.
Temperature inaccurate and cannot be calibrated.	1. Faulty connector. 2. Faulty electrode. 3. Faulty instrument.	Check the connector and replace if necessary. Fit new electrode, part number 122201. Return to factory for repair.

22. Appendices

22.1 Care, Cleaning and Maintenance of Conductivity Electrodes

22.1.1 Care of Conductivity electrodes

The conductivity section of the electrode supplied with your **WP-81** consists of two platinum wires that are plated with a layer of "platinum-black". This is quite a soft layer and is required for stable, accurate measurements. In time, the platinum-black layer may wear off in some applications, at which time the electrode will require replatinising (see section 22.2). You can help to maintain the platinum-black layer by following these simple rules:

1. **NEVER** touch or rub the electrode wires with your fingers, cloth etc.
2. Avoid using the electrode in solutions that contain a high concentration of suspended solids, such as sand or soil, which can abrade the electrode wires. Filter these types of solutions first, if possible.
3. Avoid concentrated acids. If you must measure acids, remove the electrode immediately after taking the measurement and rinse well with distilled water.

Conductivity electrodes can be stored dry. Ensure that the electrode is stored in a covered container, to avoid dust and dirt build-up.

22.1.2 Cleaning of Conductivity of Electrodes.

Platinised platinum Conductivity electrodes can only be cleaned by rinsing in a suitable solvent. **DO NOT wipe the electrode wires**, as this will remove the platinum-black layer.


1. Rinsing in distilled water will remove most build-ups of material on the electrode wires.
2. Films of oils or fats on the electrode wires can usually be removed by rinsing the electrode in methylated spirits.
3. Stubborn contamination can be removed by soaking the electrode in a solution of 1 part Concentrated HCl and 10 parts distilled water. The electrode should not be soaked for more than approximately 5 minutes, otherwise the platinum-black layer may start to dissolve.
4. If all of these methods fail, then the last resort is to physically scrub the electrode wires, which will remove the contaminant and the layer of platinum-black. Use only a cloth or nylon scouring pad. **DO NOT USE STEEL WOOL**. The electrode will then need to be cleaned in HCl, as per step 3 and replatinised, as per section 22.2.

22.2 Replatinising Conductivity Electrodes

There are several ways to replatinise Conductivity electrodes.

1. The simplest way is to return the electrode to the TPS factory. We can fully clean the electrode, replatinise it and test all aspects of its performance.
2. An automatic replatiniser is available from TPS, along with replatinising solution. This will plate the electrodes for the right amount of time at the correct current. Ordering details are as follows:

Automatic Conductivity Electrode Replatiniser	Part No 122160
20mL Platinising Solution (for approx 30 uses)	Part No 122300
3. Conductivity electrodes can be manually replatinised, according to the following procedure:
 - a) Soak the electrode in a solution of 1 part Concentrated HCl and 10 parts distilled water for approximately 5 minutes.
 - b) Rinse the electrode well in distilled water.
 - c) Immerse the electrode in platinising solution at least to the vent hole in the white plastic cover. Platinising solution is available from TPS (part no 122300).
 - d) Alternatively, platinising solution can be prepared by dissolving 1g of Hydrogen Chloroplatinate (H_2PtCl_{16}) in 30mL of distilled water, and including about 0.01g of Lead Acetate ($(CH_3COO)_2Pb$) and a drop or two of concentrated HCl.
 - e) Apply a direct current of 10mA between pins 1 and 5 of the electrode plug, as per the diagram below. Reverse the polarity every 30 seconds. After approximately 8 minutes (4 minutes per electrode wire), they should have an even "soot" like appearance. Avoid excess current and this will cause incorrect platinising.
 - f) After platinising, rinse the electrode well in distilled water.
 - g) If you have any doubts about any of these steps, then you should consider returning the electrode to the factory. The cost of replatinising is quite low, and you will be guaranteed of the best possible result.



Electrode Connector

22.3 pH Electrode Fundamentals

A combination pH Electrode is two electrodes in one. The sensing membrane is the round or spear shaped bulb at the tip of the electrode. This produces a voltage that changes with the pH of the Solution. This voltage is measured with respect to the second part of the electrode, the reference section. The reference section makes contact with the sample solution using a salt bridge, which is referred to as the reference junction. A saturated solution of KCl is used to make contact with the sample. It is vital that the KCl solution has an adequate flow rate in order to obtain stable, accurate pH measurements.

22.3.1 Asymmetry of a pH Electrode

An "ideal" pH electrode produces 0 mV output at 7.00 pH. In practice, pH electrodes, generally produce 0 mV output at slightly above or below 7.00 pH. The amount of variance from 7.00 pH is called the asymmetry. Figure 22-1 illustrates how asymmetry is expressed.




Figure 22-1

22.3.2 The Slope of a pH Electrode

As mentioned above, a pH electrode produces 0 mV output at around 7.00 pH. As the pH goes up, an “ideal” pH electrode produces -59mV/pH unit at 25 °C. As the pH goes down, an ideal pH electrode produces +59mV/pH unit. In practice, pH electrodes usually produce slightly less than this. The output of a pH electrode is expressed as a percentage of an ideal electrode. For example, an ideal electrode that produces 59mV/pH unit has “100% Slope”. An electrode that produces 50.15mV/pH unit has “85% Slope” (see Figure 22-2).




Figure 22-2

22.3.3 Temperature Compensation

The slope of a pH electrode (section 22.3.2) is affected by temperature. This effect is compensated for either by using an Automatic Temperature Compensation (ATC) probe or by entering the sample temperature manually. Figure 22-3 shows the slope of a pH electrode at various temperatures.




Figure 22-3

22.4 Checking the reference junction of a pH electrode.

If pH readings are inaccurate or unstable, the reference junction of the electrode may be blocked. The following test can be performed to determine if the reference junction of a pH electrode is making adequate contact with the sample solution.

1. Calibrate the pH section of the **WP-81**, as per section 6.1.
2. Dilute 1 part of pH6.88 buffer with 9 parts of distilled water.
3. Measure the pH of the diluted buffer. The result should be 7.06 +/-0.05 pH.
4. If the value obtained is outside of these limits, then clean the reference junction, as per the instructions supplied with the pH electrode.
5. Re-calibrate the **WP-81** and repeat the test.
6. If the value obtained is still outside 7.06 +/-0.05 pH, then the electrode should be replaced.

22.5 Determining if an instrument or pH sensor is faulty

The following test can be performed to help determine if the **WP-81** or the pH sensor is faulty.

1. Initialise the **WP-81** (see section 19).
2. Disconnect the pH electrode.
3. Connect the centre pin of the **pH** connector with the outside frame of the connector, using a short piece of wire or a paper clip etc.
4. The meter should read approximately 7.00. If you calibrate the pH readout (section 6.1), the **WP-81** will read around 6.88 pH, depending upon the temperature readout.
5. If the **WP-81** is operating correctly, the reading should be totally stable with the wire firmly in place. If not, the meter requires servicing.
6. Now carefully disconnect the wire from the centre pin only (make sure the other end of the wire remains connected to the outside frame of the connector).
7. The reading should steadily drift away from 7.00 (either up or down) at a rate of approximately 1 pH or less every 3 seconds. If the drift rate is faster than this, then input circuitry of the **WP-81** is faulty and requires servicing.

23. Warranty

TPS Pty. Ltd. guarantees all instruments and electrodes to be free from defects in material and workmanship when subjected to normal use and service. This guarantee is expressly limited to the servicing and/or adjustment of an instrument returned to the Factory, or Authorised Service Centre, freight prepaid, within twelve (12) months from the date of delivery, and to the repairing, replacing, or adjusting of parts which upon inspection are found to be defective. Warranty period on electrodes and batteries is three (3) months.

There are no express or implied warranties which extend beyond the face hereof, and TPS Pty. Ltd. is not liable for any incidental or consequential damages arising from the use or misuse of this equipment, or from interpretation of information derived from the equipment.

Shipping damage is not covered by this warranty.

PLEASE NOTE:

A guarantee card is packed with the instrument or electrode. This card must be completed at the time of purchase and the registration section returned to TPS Pty. Ltd. within 7 days. No claims will be recognised without the original guarantee card or other proof of purchase. This warranty becomes invalid if modifications or repairs are attempted by unauthorised persons, or the serial number is missing.

PROCEDURE FOR SERVICE

If you feel that this equipment is in need of repair, please re-read the manual. Sometimes, instruments are received for "repair" in perfect working order. This can occur where batteries simply require replacement or re-charging, or where the electrode simply requires cleaning or replacement.

TPS Pty. Ltd. has a fine reputation for prompt and efficient service. In just a few days, our factory service engineers and technicians will examine and repair your equipment to your full satisfaction.

To obtain this service, please follow this procedure:

Return the instrument AND ALL SENSORS to TPS freight pre-paid and insured in its original packing or suitable equivalent. INSIST on a proof of delivery receipt from the carrier for your protection in the case of shipping claims for transit loss or damage. It is your responsibility as the sender to ensure that TPS receives the unit.

Please check that the following is enclosed with your equipment:

- **Your Name and daytime phone number.**
- **Your company name, ORDER number, and return street address.**
- **A description of the fault. (Please be SPECIFIC.)**

(Note: "Please Repair" does NOT describe a fault.)

Your equipment will be repaired and returned to you by air express where possible.

For out-of-warranty units, a repair cost will be calculated from parts and labour costs. If payment is not received for the additional charges within 30 days, or if you decline to have the equipment repaired, the complete unit will be returned to you freight paid, not repaired. For full-account customers, the repair charges will be debited to your account.

- **Always describe the fault in writing.**
- **Always return the sensors with the meter.**